	Application No.	Applicant(s)	
Mating of Allowskills	10/743,253	CARR ET AL.	
Notice of Allowability	Examiner	Art Unit	<u> </u>
	Jennifer Doan	2874	
The MAILING DATE of this communication appearance All claims being allowable, PROSECUTION ON THE MERITS IS herewith (or previously mailed), a Notice of Allowance (PTOL-85) NOTICE OF ALLOWABILITY IS NOT A GRANT OF PATENT RIOF of the Office or upon petition by the applicant. See 37 CFR 1.313	(OR REMAINS) CLOSED in this ap or other appropriate communication IGHTS. This application is subject to	plication. If not included n will be mailed in due cou	ırse. THIS
1. \boxtimes This communication is responsive to <u>amendment filed on N</u>	<u>//ay 18, 2006</u> .		
2. A The allowed claim(s) is/are 1,3-13,15-25,27-29 and 31-40.			
3. ☐ Acknowledgment is made of a claim for foreign priority ur a) ☐ All b) ☐ Some* c) ☐ None of the:			
1. Certified copies of the priority documents have			
2. Certified copies of the priority documents have been received in Application No			
 Copies of the certified copies of the priority does International Bureau (PCT Rule 17.2(a)). 	cuments have been received in this	national stage application	n from the
* Certified copies not received:			
Applicant has THREE MONTHS FROM THE "MAILING DATE" noted below. Failure to timely comply will result in ABANDONN THIS THREE-MONTH PERIOD IS NOT EXTENDABLE.	of this communication to file a reply IENT of this application.	complying with the requir	rements
4. A SUBSTITUTE OATH OR DECLARATION must be subm INFORMAL PATENT APPLICATION (PTO-152) which give			ICE OF
 CORRECTED DRAWINGS (as "replacement sheets") must (a) ☐ including changes required by the Notice of Draftspers 1) ☐ hereto or 2) ☐ to Paper No./Mail Date (b) ☐ including changes required by the attached Examiner's Paper No./Mail Date 	on's Patent Drawing Review (PTO	·	
Identifying indicia such as the application number (see 37 CFR 1 each sheet. Replacement sheet(s) should be labeled as such in t			ck) of
DEPOSIT OF and/or INFORMATION about the depo attached Examiner's comment regarding REQUIREMENT	sit of BIOLOGICAL MATERIAL	must be submitted. Not	e the
Attachment(s) 1. ☐ Notice of References Cited (PTO-892)	5. Notice of Informal F	Patent Application (PTO-1	52)
2. ☐ Notice of Draftperson's Patent Drawing Review (PTO-948)	6. Interview Summary	(PTO-413),	,
3. Information Disclosure Statements (PTO-1449 or PTO/SB/0	Paper No./Mail Da		
Paper No./Mail Date 4. Examiner's Comment Regarding Requirement for Deposit	8. 🛛 Examiner's Stateme	ent of Reasons for Allowa	ince
of Biological Material	9.	Tennifer	Doan
		JENNIFEH DO PRIMARY EXAM	AN IINER

Application/Control Number: 10/743,253 Page 2

Art Unit: 2874

EXAMINER'S STATEMENT OF REASONS FOR ALLOWANCE

Response to Amendment

 Applicant's amendment filed on May 18, 2006 has been fully considered and entered.

Reasons for Allowance

- 2. Claims 1, 3-13, 15-25, 27-29 and 31-40 are allowed.
- 3. The following is an examiner's statement of reasons for allowance:

The prior art of record fails to disclose or reasonably suggest a method and a device comprising an optical waveguide having a first grating, wherein at least a portion of waveguide has a functional layer adapted to bind an analyte; a plurality of grooves in the optical waveguide from the first grating; at least some of the grooves are covered by the functional layer; further a device comprising a MZI having two arms, wherein one arm has an optical resonator which has a functional layer adapted to bind an analyte and further wherein one arm of the MZI includes the optical waveguide having the first and second gratings form an optical resonator in combination with the other limitations of claims 1, 12, 22, 24, 25, 29, 32 and 35.

Claims 3-11 depend from claim 1.

Claims 13, 15-21 depend from claim 12.

Claims 23 and 38 depend from claim 22.

Application/Control Number: 10/743,253

Art Unit: 2874

Claims 27 and 28 depend from claim 25.

Claim 31 depends from claim 29.

Claims 33 and 34 depend from claim 32.

Claims 36 and 37 depend from claim 35.

The prior art of record also fails to disclose or reasonably suggest a method and a device comprising an optical waveguide having a first grating, wherein at least a portion of waveguide has a functional layer adapted to bind an analyte; the device further comprises one or more other optical waveguides, each having a grating, wherein at least two optical waveguides have different functional layer adapted to bind different analytes in combination with the other limitations of claims 39 and 40.

Page 3

Any comments considered necessary by applicant must be submitted no later than the payment of the issue fee and, to avoid processing delays, should preferably accompany the issue fee. Such submissions should be clearly labeled "Comments on Statement of Reasons for Allowance."

Conclusion

4. Any inquiry concerning this communication or earlier communications from the examiner should be directed to Jennifer Doan whose telephone number is (571) 272-2346. The examiner can normally be reached on Monday to Thursday from 6:00am to 3:30pm, second Friday off.

Application/Control Number: 10/743,253

Art Unit: 2874

5. If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Rodney Bovernick can be reached on (571) 272-2344. The fax phone

number for the organization where this application or proceeding is assigned is 703-

The most for the organization where this application of proceeding is assigned is

872-9306.

6. Information regarding the status of an application may be obtained from the

Patent Application Information Retrieval (PAIR) system. Status information for

published applications may be obtained from either Private PAIR or Public PAIR.

Status information for unpublished applications is available through Private PAIR only.

For more information about the PAIR system, see http://pair-direct.uspto.gov. Should

you have questions on access to the Private PAIR system, contact the Electronic

Business Center (EBC) at 866-217-9197 (toll-free).

JENNIFER DOAN
PRIMARY EXAMINER

Page 4

JD

June 1, 2006